

## LABORATORY EQUIPMENT FOR SCHOOLS

Package No. <sup>(1)</sup>	Item Serial Number	Item Name	Qty
		<b>Burette 50ml</b>	

**Item Description**

A uniform-bore glass tube with fine gradations and a stopcock at the bottom, used especially in laboratory procedures for accurate fluid dispensing and measurement. The product should meet international quality standard.

**Image**



(Sample Required/Not Required) <sup>(2)</sup>

Specifications	Conformity		Variations (if any)
	Yes	No	
Borosilicate glass			
Interchangeable stop cock giving perfect flow control			
Resistant to chemical attack			
Permanent graduation			
Capacity 50 ml			
Graduation 0.1 ml			

Country of origin .....

Catalogue Reference (if any) .....

Country of manufacture .....

Whether the sample is submitted or not? <sup>(3)</sup>

Yes		No	
-----	--	----	--

Name and address of the Bidder.....  
 .....  
 .....  
 .....

Signature of the Bidder  
 Date .....

Rubber stamp

(1) - Package No. under the Tender

(2) -The Buyer shall indicate whether a sample of the item is required or not, at the tendering stage.

(3) -The Bidder shall note that whenever a sample is required <sup>(2)</sup>, the Tender will be considered for evaluation only on approval of the sample.