

LABORATORY EQUIPMENT FOR SCHOOLS

Package ⁽¹⁾ No.	Item Serial Number	Item Name	Qty
		Ammeter (0 – 05 A) DC	

Item Description

The Ammeter is mounted on bakelite base with transparent sides and top. All parts of the movement are visible. Complete with two terminals, black & red and a plastic case. The product should meet international quality standard.

Image



(Sample Required/Not Required)⁽²⁾

Specifications	Conformity		Variations (if any)
	Yes	No	
Moving coil overload protected DC Ammeter			
Large clear dial with anti parallax mirror			
Accuracy $\pm 2.5\%$ or better			
External zero-adjust calibration			
Sturdy protective transparent cover			
Scale 0 - 5A x 0.1A.			
Red & Black leads			
Scale not less than 50 mm long			

Country of origin

Catalogue Reference (if any)

Country of manufacture

Whether the sample is submitted or not?⁽³⁾

Yes		No	
-----	--	----	--

Name and address of the Bidder.....

.....
 Signature of the Bidder
 Date

Rubber stamp

(1) - Package No. under the Tender

(2) -The Buyer shall indicate whether a sample of the item is required or not, at the tendering stage.

(3) -The Bidder shall note that whenever a sample is required⁽²⁾, the Tender will be considered for evaluation only on approval of the sample.