

LABORATORY EQUIPMENT FOR SCHOOLS

Package ⁽¹⁾ No.	Item Serial Number	Item Name	Qty
		Ammeter (0 – 100 Micro A) DC	

Item Description

The Ammeter is mounted on bakelite base with transparent sides and top. All parts of the movement are visible. Complete with two terminals, black & red and a plastic case. The product should meet international quality standard.

Image



(Sample Required/Not Required) ⁽²⁾

Specifications	Conformity		Variations (if any)
	Yes	No	
Moving coil, overload protected DC Micro ammeter			
Large clear dial with anti parallax mirror			
External zero-adjust calibration			
Accuracy $\pm 2.5\%$ or better			
Scale 0 -100 mA x 2 mA, not less than 50 mm long.			
Resistance with diode protection			
Red & Black leads			
Transparent sturdy protective cover			

Country of origin

Catalogue Reference (if any)

Country of manufacture

Whether the sample is submitted or not? ⁽³⁾

Yes		No	
-----	--	----	--

Name and address of the Bidder.....

.....

.....
 Signature of the Bidder

Date

Rubber stamp

(1) - Package No. under the Tender

(2) -The Buyer shall indicate whether a sample of the item is required or not, at the tendering stage.

(3) -The Bidder shall note that whenever a sample is required ⁽²⁾, the Tender will be considered for evaluation only on approval of the sample.