

## LABORATORY EQUIPMENT FOR SCHOOLS

Package <sup>(1)</sup> No.	Item Serial Number	Item Name	Qty
		<b>Barometer Aneroid</b>	

**Item Description**

Barometer Aneroid is a device used to measure atmospheric pressure. The product should meet international quality standard.

**Image**



(Sample Required/Not Required) <sup>(2)</sup>

Specifications	Conformity		Variations (if any)
	Yes	No	
An accurately reading aneroid barometer with ~ 150mm diameter.			
Silvered dial			
Mounted in a sturdy protective transparent case with provision for wall mounting.			
Stout cover glass, index pointer, setting knob.			
Graduation 900 mb – 1100 mb x 2mb.			
Accuracy ± 1.0 mb or better			

Country of origin .....

Catalogue Reference (if any) .....

Country of manufacture .....

Whether the sample is submitted or not? <sup>(3)</sup>

Yes	<input type="checkbox"/>	No	<input type="checkbox"/>
-----	--------------------------	----	--------------------------

Name and address of the Bidder.....  
 .....  
 .....  
 .....

.....  
 Signature of the Bidder  
 Date .....

Rubber stamp

(1) - Package No. under the Tender

(2) -The Buyer shall indicate whether a sample of the item is required or not, at the tendering stage.

(3) -The Bidder shall note that whenever a sample is required <sup>(2)</sup>, the Tender will be considered for evaluation only on approval of the sample.